

**BIDDING FOR THE PROCUREMENT OF VARIOUS  
DENTAL EQUIPMENT**

**BID BULLETIN NO. DE 2019-001**

This bid bulletin clarifies queries sent through e-mail and raised during the Pre-Bid Conference held last 15 October 2019 including other matters relative to the **Bidding for the Procurement of Various Dental Equipment**.

**1. Queries/Questions by Prospective Bidders**

| <b>Issues Raised</b>  | <b>Clarifications</b>  |
|---|--|
| 1. On the amount of Bidding Documents.  | The Bidding Documents can be purchased in the amount of PhP5,000.  |
| 2. On the required specifications of Bench Grinder.   | Below are the required specifications based on the TOR for BENCH GRINDER: <ul style="list-style-type: none"> <li>- Power House Bench Grinder 10"</li> <li>- Model PH-250 Flexible Shaft</li> <li>- 750 watts, one (1) horse power (hp)</li> </ul>  |
| 3. On the required specifications of Oxygen Acetylene Tank.   | Below are the required specifications based on the TOR for OXYGEN ACETYLENE TANK: <ul style="list-style-type: none"> <li>- Size: 10cu.ft</li> <li>- Weight: 8.75 lbs.</li> </ul>   |
| 4. On the required number of years in existence of the prospective bidder/supplier of Dental Equipment.   | The prospective bidder/supplier of Dental Equipment must have at least 3 years of existence.   |
| 5. On the required Delivery Schedule.   | The winning bidder must deliver the Dental Equipment within sixty (60) calendar days upon the issuance of the Purchase Order.  |
| 6. Whether the prospective bidders should submit Certificate/s that they will provide reliable and prompt after sales services from their previous clients. | Based on the TOR, the winning bidders must be able to provide a reliable and prompt after sales services. In this regard, bidders will issue a Certificate that they will provide reliable and prompt after sales services, <u>if awarded the Contract for Various Dental Equipment.</u><br><br>The prospective bidders are not required to provide a certificate that they have provided reliable and prompt after sales from their previous clients. |

|  |   |
|--|---|
| 7. On the required Warranty Service for the Dental Equipment.  | All items shall have at least one (1) year warranty upon issuance of Certificate of Acceptance.   |
| 8. On the required number of days for the Replacement of the Dental Equipment.                         | The required no. of days for the replacement of dental equipment shall be 60 days from BCDA's request.  |
| 9. Whether the end-user will provide a Certificate of Acceptance upon receipt of the Dental Equipment. | The end-user will provide a Certificate of Acceptance upon receipt of <b>all</b> the Dental Equipment.  |
| 10. On the time of return of the Retention Money to the winning bidder.                                | The Retention Money shall be returned to the winning bidder upon the issuance of the Certificate of Satisfactory Delivery.  |
| 11. On the issuance of Certificate of Satisfactory Delivery.   | The end-user shall issue a Certificate of Satisfactory Delivery 90 days after the issuance of the Certificate of Acceptance provided that the Winning Bidder has no pending deliverables. |
| 12. Whether the bidders can request for the soft editable copies of the sample forms.                  | Except for the Technical Compliance form, the bidders may be provided with the editable soft copies of the sample forms upon purchase of the bidding documents.                           |

**The clarifications herein issued amend, modify or reise any provision in the TOR which are inconsistent or contradictory thereto.**

## **2. Clarifications from the Bids and Awards Committee (BAC) for Goods**

- a. The bid should not be more than the ABC, otherwise the bidder will be disqualified.
- b. The bid documents can be accessed from the PhilGEPS and BCDA websites.
- c. Statement of Bidder of all on-going and completed government or private contracts:
  - (SF-G&S-19A) Statement of all its ongoing government and private contracts, including contracts awarded but not yet started, if any, whether similar or not similar in nature and complexity to the contract to be bid; and
  - (SF-G&S-19B) Statement of Single Largest Completed Contract (SLCC) similar in nature to the contract to be bid equivalent to at least fifty percent (50%) of the ABC or two (2) similar contracts with an aggregate amount equivalent to at least fifty percent (50%) of the ABC within ten (10) years from the date of submission and receipt of bids.

The two (2) statements required shall indicate for each contract the following:

- name of the contract;
- date of the contract;
- contract duration;
- owner's name and address;
- kinds of Goods;
- For Statement of Ongoing Contracts - amount of contract and value of outstanding contracts;
- For Statement of SLCC - amount of completed contracts, adjusted by the Bidder to current prices using PSA's consumer price index, if necessary for the purpose of meeting the SLCC requirement;
- date of delivery; and
- end user's acceptance or official receipt(s) or sales invoice issued for the contract, if completed, which shall be attached to the statements.

**Notes:**

- Statement of all ongoing contracts requirement is merely a "statement" of all ongoing contracts. It refers to a "statement" to be made by the bidder relative to all its ongoing government and private contracts, and not the actual submission of the physical NOA, NTP or Contracts.
  - The Statement of SLCC to be submitted by the bidder shall attach **ANY** of the following proof:
    - end user's acceptance; or
    - official receipt of the last payment received; or
    - sales invoice
  - One (1) completed contract of similar nature must have a value of at least 50% or two (2) similar contracts with an aggregate amount equivalent to at least fifty percent (50%) of the Approved Budget for the Contract (ABC) of bidding at hand.
- d. The necessary bid bulletin shall be posted in PhilGEPS and BCDA website on **18 October 2019 (Fri)**. However, only those who purchased the bidding documents are entitled to directly receive a copy of the Bid Bulletin by email.
- e. Only those bidders who purchased the bid documents are allowed to submit their bids.
- f. The computation of the bid security shall be based on the Approved Budget for the Contract (ABC).

- g. The bidders may submit a notarized Bid Securing Declaration in lieu of other forms of bid security. It is important to note, however, that although the Bid Securing Declaration does not require cash, its enforcement shall include automatic blacklisting and payment of certain fines in the event that the bidder violates any of the conditions set forth in the bidding documents.
- h. NFCC Computation shall be based on the audited Financial Statements for the year 2018, using the computation provided in paragraph 5.5 of the Instruction to Bidders, and should be at least equivalent to the ABC.
- i. The bidders are requested to use tabs in compiling their bid documents for quick and easy identification and verification.

#### 4. Reminders

- a. BCDA adheres to the "No-Contact Rule". All clarifications shall be made in writing and addressed to the BAC-G Secretariat. **Deadline for Clarification is on 17 October 2019 (Thur) at 5:00PM.**
- b. Visiting or calling the members of the BAC, the TWG, the Secretariat or anyone working for BCDA is not allowed and will not be entertained.
- c. Dates, in the absence of any qualifications, are meant to be calendar days. Calendar days include Saturdays, Sundays, and Holidays.
- d. **The deadline for the submission of bids for the Bidding for the Procurement of Various Dental Equipment is on 25 October 2019 (Fri) at 10:30 AM** at the BCDA Central Receiving and Releasing Area (CRRA) located at the 2<sup>nd</sup> Floor Bonifacio Technology Center, 31<sup>st</sup> St. cor. 2<sup>nd</sup> Avenue, Bonifacio Global City, Taguig City. The computer system clock at the CRRA that is set to Philippine Standard Time (PhST) shall be used as reference in determining the time of the submission of bids. Hence, participating bidders are advised to synchronize their timepieces with the said computer system clock. **Late bids or those submitted after 10:30 AM of 25 October 2019 (Fri) shall not be accepted.**
- e. Bidders may submit their eligibility documents days ahead of the deadline for submission in order to avoid late submission.
- f. **Bid opening shall be on 25 October 2019 (Fri) at 11:30 AM** at the BCDA Corporate Center, 2<sup>nd</sup> Floor Bonifacio Technology Center, 31<sup>st</sup> St. cor. 2<sup>nd</sup> Avenue, Bonifacio Global City, Taguig City. Bids will be opened in the presence of the Bidders. Although the bidders' attendance is not compulsory during the Opening of Bids, it is advised that bidders send their representative to assist the BAC and answer clarifications, if there is any.

- g. Each and every page of the Bid Forms must be appropriately signed by the bidder or the bidder's authorized representative. The authorization should also be attached. Failure to do so shall be a ground for rejection of the Bid.
- h. All licenses, permits, and other required clearances should be valid at the time of the submission of bids, Post-Qualification Evaluation, and signing of the contract.
- i. The BAC expects the bidders to exercise due diligence in going through the bidding documents so that they can prepare their bids intelligently.
- j. The following documents are required to be NOTARIZED, which should be duly sworn to before a Notary Public, and include complete information on the PTR No., MCLE No., Authority or Commission Number, etc. of the Notary Public:
  - i. Omnibus Sworn Statement
  - ii. Bidder's Representative Authorization (as provided in Paragraph 2 of the Omnibus Sworn Statement )
  - iii. Bid Securing Declaration
- k. BCDA reserves the right to waive minor defects in forms and requirement as long as they do not affect the genuineness and authenticity of the documents submitted.
- l. BCDA reserves the right to accept or reject any bid, to annul the bidding process, and to reject all bids at any time prior to contract award, without thereby incurring any liability to the affected bidder or bidders.

**Please be reminded that all queries after the release of this Bid Bulletin will not be entertained.**

All provisions, conditions and statements in the bidding documents inconsistent with this Bid Bulletin are either modified or rendered ineffective, as the case may be.

**Issued on 18 October 2019.**

#### **BIDS AND AWARDS COMMITTEE FOR GOODS**

  
**ATTY. MARIA SOLEDAD C. SAN PABLO**  
Vice Chairperson

*Bid Bulletin No. DE-2019-001*  
*October 2019*

**BCDA**  
Bases Conversion and  
Development Authority

BAC for Goods  
  
**BACG2019 – 0179**

Page 5 of 5

## TECHNICAL SPECIFICATIONS COMPLIANCE FORM

Bidders must indicate whether the merchandise offered is “Compliant” or “Non-Compliant” to the corresponding specifications prescribed by BCDA using this form.

| No. | Technical Specifications  | STATEMENT OF COMPLIANCE |               |
|-----|---|-------------------------|---------------|
|     |   | Check Appropriate Box   |               |
|     |   | Compliant               | Non-compliant |
|     | <b>DENTAL ALLOY GRINDER (2 UNITS):</b> <ul style="list-style-type: none"> <li>● ¼ Horse Power (hp)</li> <li>● 25,000 rpm</li> <li>● Weight 17 kg.</li> </ul>  |                         |               |
|     | <b>BENCH GRINDER (2 UNITS):</b> <ul style="list-style-type: none"> <li>● Power House Bench Grinder 10”</li> <li>● Model PH-250 Flexible Shaft</li> <li>● 750 watts, one (1) horse power (hp)</li> </ul>   |                         |               |
|     | <b>PORCELAIN FURNACE (1 UNIT):</b> <ul style="list-style-type: none"> <li>● Max Temperature 1200 Celsius or minus 1 degree Celsius</li> <li>● Dimension: 245mm (w) x 425mm (h) x 410mm (d), Weight 16 kg.</li> <li>● Muffle Chamber, 120x90mm Indicator</li> <li>● Must have LED indicators for furnace alerts (Green- ready, Yellow- dry, Red- heat, Blue- cool down)</li> <li>● Power requirement: 230 volts/ 50 hz.</li> </ul> |                         |               |
|     | <b>VALPLAST FLEXIBLE DENTURE INJECTION SYSTEM (2 UNITS)</b> <ul style="list-style-type: none"> <li>● Heating Temperature variable includes 2 bantam flask</li> <li>● Dimensions: 77cm x 55cm x 55cm</li> <li>● 2 copper gasket, 2 fuse</li> <li>● 45 kg. Weight</li> <li>● Silicon Oil</li> <li>● 400 watts, 220 volts/60hz</li> </ul>  |                         |               |
|     | <b>DIGITAL POLISHING UNIT ( 2 UNITS)</b> <ul style="list-style-type: none"> <li>● Weight 2.8kg, 220v-240v/50Hz</li> <li>● 8 x 9.5 x 5.25 inches</li> <li>● 100 watts</li> </ul>   |                         |               |

|  |  |  |  |
|--|--|--|--|
|  | <ul style="list-style-type: none"> <li>• With positive and negative terminal</li> <li>• Electrolytic Clamp and plate</li> <li>• With digital display</li> <li>• Amp and Time Adjust Volume</li> <li>• With air pump control, memory control, run control</li> </ul>  |  |  |
|  | <p><b>CENTRIFUGAL CASTING MACHINE (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Heavy duty spring and adjustable balance mechanism<br/>Weight 13lbs.</li> <li>• Dimension: 16.2 x 9.2 x 9.1</li> <li>• 1-year warranty</li> </ul>   |  |  |
|  | <p><b>BIO ART SEMI ADJUSTABLE ARTICULATOR (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Arcon Type; curved condylar guide</li> <li>• Adjustable condylar guide and bennet angle</li> <li>• Central lock; stabilization system of condylar guide</li> </ul>  |  |  |
|  | <p><b>PATTERSON TRIMMER (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• 12 x ¼ model trimmer coarse abrasive wheel</li> <li>• Equipped with fully adjustable table</li> <li>• Dimension: 16' wide x 17' high x 15'' deep</li> <li>• Coarse abrasive wheel (40.5cm x 43cm x 38cm, 0 to 1.5mm for each side</li> <li>• 19.3 kg (weight), ½ horse power</li> <li>• 230volts / 50hz / 4amp</li> </ul> |  |  |
|  | <p><b>DENTAL VACUUM FORMING MACHINE (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Symmetrical heating system with uniform heat distribution</li> <li>• Adjustable rust resistant body</li> <li>• Compatible with all dental thermoplastics</li> <li>• 1000watts, 220 volts</li> <li>• Dimension: 25cm x 21cm x 30cm</li> <li>• 6kg. weight</li> </ul>   |  |  |
|  | <p><b>DENTAL SURVEYOR (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Complete standard gauge tips</li> <li>• Aluminum electrostatic paint finish</li> </ul>  |  |  |

|  |  |  |  |
|--|--|--|--|
|  | <ul style="list-style-type: none"> <li>• 1 swivel arm and central articulated joint</li> </ul>   |  |  |
|  | <p><b>ENDODONTIC MACHINE (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Programmable Memory: 3</li> <li>• Power Source: Battery/AC operated</li> <li>• Dimension: 92mm (w) x 148mm (d) x 124mm (h)</li> </ul>  |  |  |
|  | <p><b>APEX LOCATOR (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Battery: Continuous operation for 60 hours</li> <li>• Display: 3 color LED indicating of file movement inside the canal</li> <li>• Power Supply: 1.5volts, AAA (LR03 or R03 batteries)</li> <li>• 76kg weight (without batteries)</li> </ul> |  |  |
|  | <p><b>GAS WELDING KIT (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Complete set and accessories</li> </ul>   |  |  |
|  | <p><b>OXYGEN ACETYLENE TANK (2 UNITS)</b></p> <ul style="list-style-type: none"> <li>• Size: 10cu.ft</li> <li>• Weight: 8.75 lbs.</li> </ul>   |  |  |
|  | <p><b>Warranty Period: One (1) Year</b></p>  |  |  |

**Bidder's Authorized Representative:**

\_\_\_\_\_  
Signature over Printed Name

\_\_\_\_\_  
Principal Bidder / Supplier